

Performance of AxSYM HIV Ag/Ab Combo assay, a 4th generation assay for routine HIV screening

> Christiane Claessens, Claire Béliveau HIV Testing Conference, March 25, 2010

> > Institut national de santé publique Québec * *

www.inspq.qc.ca

Québec and HIV

Population

- 8,000,000 (23 % of Canada)
- Persons living with HIV/AIDS (2008)
- 14,000-21,300 (Canada 65,000)

Annual new cases

- 500-1,200 (Canada 2,300-4,300)
- Annual HIV tests
 - ≈ 300,000



CANAD

HIV testing in Québec

33 testing sites

- 25 hospital laboratories
- 8 private laboratories
- In 2007, 25/33 laboratories used
- AxSYM HIV-1/2 gO (Abbott), a 3rd gen. assay
 LSPQ
- Supplemental tests on all reactive specimens
 - > WB, p24 Ag EIA, HIV-2 EIA, HIV-1/HIV-2 LIA
 - On request, if acute infection suspected and Abs negative
 - p24 Ag EIA



Introduction of a 4th generation assay

Fall 2007

- The first 4th generation assay was licensed for diagnosis in Canada:
 - AxSYM HIV Ag/Ab Combo* (Abbott)

*not available in the US



Assays format

AxSYM HIV-1/HIV2 gO

3rd gen assay

- microparticle immunoassay
 - recombinant Ag
 - HIV-1 M and group O Env
 - HIV-1 core
 - HIV-2 Env
 - synthetic peptides
 - HIV-1 Env
 - HIV-2 Env.

AxSYM Ag/Ab HIV Combo 4th gen assay

- microparticle immunoassay
 - recombinant Ag
 - HIV-1 M and group O Env
 - HIV-1 core
 - HIV-2 Env
 - synthetic peptides
 - HIV-1 Env
 - HIV-2 Env.
 - HIV p24 monoclonal Ab

Detection of fluorescent products



Detection of fluorescent products

Sensitivity assessment - Ag/Ab assay

Sera from patients with documented acute infection :

24 non reactive with 3rd gen test:

- 23 Ag POS
- 1 Ag NEG (3 months before WB POS)


32 reactive (signal/cutoff values <10.0) with 3rd gen test:

32 Ag POS (15 WB NEG; 17 WB IND)

Results

All p24 positive samples were reactive with 4th gen assay





Introduction of Ag/Ab in routine use

In 2008, sites using AxSYM platform switched to the Ag/Ab : Algorithm adopted:

- S/CO < 0.90 \rightarrow no further testing
 - Negative for Ab and Ag
 - $0.90 \le S/CO < 1.00 \rightarrow Greyzone (GZ) \rightarrow repeat$
 - If GZ on repeat testing:
 - Equivocal, follow-up specimen (3 6 weeks) requested
 - S/CO > 1.00 on repeat testing
 - Referred to LSPQ for supplemental testing



Objectives

Assess the impact of the introduction of this new Ag/Ab test on

- the specificity of HIV serological results
- the detection of acute infections.



Methods

Data examined:

 Trends in reactivity and confirmatory rates before and after implementation of the Ag/Ab test

- Cases of acute infection detected
- Limitation of the study:

Sequential data, no parallel testing of specimens



Trends in reactive and HIV confirmed specimens

	2007 3 rd gen	2008 Ag/Ab
Tested	294 308	300 923
Reactive	0.78 % <mark>(2 290)</mark>	0.74 % <mark>(2 228)</mark>
Confirmed	70.8 % (1 622)	71.6% <mark>(1 595)</mark>

de santé publique Québec 🏘 🛤

Trends in selected sites

	Site A		Site B		Site C	
	3 rd	Ag/Ab	3 rd	Ag/Ab	3 rd	Ag/Ab
Tested	30,530	32,193	29,773	31,226	47,230	48,688
Reactive	<mark>371</mark> (1.2%)	<mark>350</mark> (1.1%)	147 (0.49%)	145 (0.46%)	229 (0.48%)	226 (0.46%)
Confirmed	304 (81.9%)	332 (94.8%)	74 (50.3%)	<mark>92</mark> (63.4%)	126 (55.0%)	144 (63.7%)



Trends in false positive p24 EIA tests

	2007	2008	
	3 rd gen	Ag/Ab	
p24 Ag EIA	10	9	
(NT <50%)	(0.0034 %)	(0.0030 %)	



Detection of acute HIV infection

	2007	2008
	3 rd gen	Ag/Ab
Ab (WB) Neg /Ag Pos	12	22
WB Ind /Ag Pos	22	29



Acute HIV infection

Case 644120

53 y. old male presenting with AHI symptoms , ART initiated on day 16

Day	Ag/Ab	WB	p24 Ag	LIA
0	52.2	NEG	POS	
10	6.9	IND p24+,p18+	POS	
26	6.1	IND gp160±, 55±,p40+,p24++,p18++	NEG	POS-1
42	4.8	IND gp160± p55± p40+ p24++ p18±	NEG	POS-1



Observations

In cases of acute HIV infection we noticed:


- Wide range of S/CO values (1.02 52.2) for WB negative/Ag p24 positive specimens,
- A significant decrease in the S/CO values between sera obtained days apart.



Transient reduction of reactivity

Case 626973


Transition zone between decreasing levels of p24 Ag and increasing levels of Abs



Transient reduction of reactivity

Case 637587

Transition zone between decreasing levels of p24 Ag and increasing levels of Abs



Window period exists...

CASE 618350

Day	Ag/Ab	WB	p24
0	< 0.9		
10	6.4	NEG	POS
38	18.6	POS	



but is reduced compared to 3rd gen

CASE 637587

Day	Ag/Ab	3 rd gen	WB	p24 Ag
0	1.87		NEG	POS
1	5.6		NEG	POS
7		1.39	NEG	POS
9	24.8		NEG	POS
36	9.13		POS	



Conclusion

Implementation of 4th generation assay in routine use in the Québec Provincial program :

- Allowed early identification of individuals with acute infection
 - Did not result in more false-positives



Discussion

Ag/Ab detected more patients with acute infection.

"One size fits all" approach is simple for detection of acute infection, no need for specific request of p24 Ag test.

Confirmatory algorithm must be reviewed.

Tests to detect HIV RNA or p24 Ag are needed to clarify the status of a Ag/Ab reactive/WB negative sample.

Window period is reduced.

- A second window period can occur.
- Follow-up is suggested for all greyzone reactors.



Note to laboratory personnel...

With Ag/Ab assays, sample mismatch is no longer the most probable explanation for a strongly reactive, WB negative specimen.



Acknowledgments

• Members of *Programme provincial de diagnostic de laboratoire du VIH*

Abbott Laboratories provided Ag/Ab kits for sensitivity assessment

